



CERTIFICATE OF ANALYSIS

Product Information

Part Number	1N38C2
Product Name	Rapid 2019-nCoV IgG/IgM Combo Test Card
Lot Number	20031010
Quantity	350 Tests
Exp. Date	2021-09
Integrity Test	Pass

Assay Result

Controls	No. Tested	No. Passed	No. Failed
Negative	10	10	0
IgG Weak Positive	3	3	0
IgM Weak Positive	3	3	0
Positive	4	4	0

The lot (check one) passes fails QC test

Authorized signature: Liu Bin

Date: 2020-03-10